

EC – DECLARATION OF CONFORMITY

In compliance with Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.

COMFORTIES EXAMINATION GLOVES, LATEX, POWDER-FREE

Page 1 of 1

REF	Size	Intended use	K = MP Class R = Rule	UMDNS
8900	XS	Medical gloves for single use, intended to protect both the wearer and the patient from cross-contamination caused by micro-organisms. Made of natural rubber latex. Protein content: < 50 µg/g	K1-R5	11-882
9000	S		K1-R5	11-882
9100	M		K1-R5	11-882
9200	L		K1-R5	11-882
9300	XL		K1-R5	11-882

Harmonised standards: EN 455 1, 2, 3, 4

As a manufacturer we declare under our responsibility that these medical aids meet the requirements of the above-mentioned Directive.
In accordance with that Directive, Annex VII, the products bear the CE-marking.

Version: 1
Date: August 2017
Valid until: July 2022

Name: Georgi Kostadinov
Department: Regulatory Affairs
Signature :



Headquarters:
Comforties Ltd.
Liuben Karavelov Str. 6
9000 Varna

comforties.com UK
219 Kensington High Street,
Kensington, London
W8 6BD

comforties.com IE
Kandoy House,
2 Fairview Strand
Dublin 3

T +44 203 608 76 06 (UK)
+353 15 133338 (IE)
E en.info@comforties.com (UK)
ie.info@comforties.com (IE)

comforties.com